


Certificate of Analysis

Product Name:

Litmus Milk - suitable for microbiology, NutriSelect® Plus

Product Number: 17158
Batch Number: BCCN5016
Brand: SIAL
Quality Release Date: 02 JUL 2025
Expiration Date: MAY 2030

Test	Specification	Result
Appearance (Color) Faint Blue and Faint Purple and Faint Red	Conforms to Requirements	Faint Blue
Appearance (Form)	Powder	Powder
Microbiological Test Pass as Described in Data Sheet	Pass	Pass
Solubility (Color)	Faint Purple to Purple	Purple
Solubility (Turbidity) 101 g/l in Water, 121 C for 5 min.	Clear to Turbid	Turbid
pH at 25 C	6.6 - 7.0	7.0



Dr.Reinhold Schwenninger
 Quality Assurance
 Buchs,Switzerland CH

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

